




Dynamic Particle Detector R&D at Chung-Ang University

Position-sensitive scintillation detectors

CPNR Workshop 2025
Chung-Ang University
Chang Hyon Ha


Optically Dynamic: Layered Detector Concept Plastic Scinitillator-Crystal (Water)-Liquid Scintillator-Acrylic





- Cherenkov Radiation:
 Direction, Threshold (PID), and Low attenuation.
- Scintillation : High Light, Low threshold, and Neutron capture
- Segmentation: Position

Assume LSC and inner tank is covered with PMTs

Triple phosphor sandwich (Phoswich) detector








Crystal signal and Organic scintillator signal are separated by a single readout


Plastic vs Liquid Scintillator (Phoswich)



Plastic and Liquid scintillator signals are separated by using machine learning

Small scale R&D (On-going) Asymmetry PS O(PMT $Q(PMT_{bottom})$

Chemically Dynamic: Isotope loading and extracting concept to LS




Motivation

How do we go about the kTon-scale next generation $0\nu\beta\beta$ detector?

- LSC space ($\phi=20m, H=20m$, ~6kTon liquid, Yemilab) is ready —> It will eventually come true. The liquid would either be water or LS
- Metal loading technology (~% level) is widely studied nowadays.
- Roughly 1% metal loading gives 60 tons of which N.A. is on average 10% (6 ton targets).
 - C.f.: 320 kg of (enriched) ^{136}Xe target on LS(decane+PC) by KamLAND-Zen
 - C.f.: ~hundreds of ^{130}Te target on LAB by SNO+
- Korea has LS expertise and underground space.
- Isotopes above 2614 keV, quality LS, and high concentration target mass would be competitive.


Current R&Ds and Sensitivity




1% nat. Ca loading would give us 114kg * 5yr Exposure (bkg. better, mass lower) For Cd, 22,500 kg yr exposure (higher mass, a bit of bkg will not be competitive), For Se and Mo, somewhere in between.

Geometrically Dynamic: Change in dynamic range. Fishing-line style sensor array


Baseline Mode




- Total Energy mainly measured by PMTs on the detector surface
- LS-embedded sensors (SiPMs or small PMTs) intercept fraction of light
- Lines can be moveable and reconfigured easily.




Geometrically Dynamic: Change in dynamic range. Fishing-line style sensor array




- Total Energy mainly measured by PMTs on the detector surface
- LS-embedded sensors (SiPMs or small PMTs) intercept fraction of light
- Lines can be moveable and reconfigured easily.




Geometrically Dynamic: Change in dynamic range. Fishing-line style sensor array




- Total Energy mainly measured by PMTs on the detector surface
- LS-embedded sensors (SiPMs or small PMTs) intercept fraction of light
- Lines can be moveable and reconfigured easily.



Geometrically Dynamic: Change in dynamic range. Fishing-line style sensor array




- Total Energy mainly measured by PMTs on the detector surface
- LS-embedded sensors (SiPMs or small PMTs) intercept fraction of light
- Lines can be moveable and reconfigured easily.





Journey to Position-sensitive detector R&Ds


Plastic scintillator array + SiPM readout system


























Plastic Scintillator Panels with Optical Fibers




Plastic Scintillator Muon Detector v4

Study of cosmic-ray muon trajectories








- PS panel (490mm x 600mm) + 28 WLS fibers + 14 SiPMs
- Calculate the center of gravity of charge from each SiPM to reconstruct the muon incident position

Position-resolution estimates


- Two PS pieces (PS : Fermilab sample)
- The dimension of trigger detector: 62.5mm x 51.2 mm




- Moving the trigger detector along X or Y axis
- Only the trigger detector participate in trigger




Muon Event Selection






- Different color with different real position of the trigger detector
- Resolution of Y-axis = 24.20 ± 0.17 mm (calculated with only three point)
- Resolution of X-axis = 38.07 ± 0.19 mm



- · Different color with different real position of the trigger detector
- Resolution of Y-axis = 24.20 ± 0.17 mm
- Resolution of X-axis = 38.07 ± 0.19 mm

22

Muon Tomography



- Simple and compact Cosmic Ray muon detector using SiPMs
- Easily Portable Detector with custom DAQ under 68 Watt. power comsumption
- Rugged detector with a good PID
- Wide Area Mapping in short measurement duration.
- Position-sensitive detector.
- Educational Demonstration (Real-time data displays)



Radiation Physics and Chemistry 225 (2024) 112117



Contents lists available at ScienceDirect

Radiation Physics and Chemistry


journal homepage: www.elsevier.com/locate/radphyschem





Real-time portable muography with Hankuk Atmospheric-muon Wide Landscaping : HAWL

J. Seo a, N. Carlin b, D.F.F.S. Cavalcante b, J.S. Chung a, L.E. França b, C. Ha a, J. Kim a, J.Y. Kim a, H. Kimku a, B.C. Koh a, Y.J. Lee a, B.B. Manzato b, S.W. Oh a, R.L.C. Pitta b, S.J. Won a



^a Department of Physics, Chung-Ang University, Seoul, 06974, Republic of Korea

b Physics Institute, University of São Paulo, São Paulo, 05508-090, Brazil


Hankuk Atmospheric-muon Wide Landscaping (HAWL)





Hankuk Atmospheric-muon Wide Landscaping (HAWL)




All highway tunnels (47 structures) are mapped within two hours









Transition towards no fibers and SiPM array



250 300 Charge(ADC)

HAWL Upgrade





Charge Spectrum (x4 increase)





- Large SiPM coverage $(1mm^2 \rightarrow 36mm^2)$
- Direct attachment to the scintillator (No fibers)
- Position Sensitivity

 Boryeung Undersea Tunnel (Earthquake warning)

Upper panel is being manufactured.

- Subway (Sink hole)
- Ulleungdo Lava bed (Volcano)





HAWL Upgrade Trip: Boryoung Tunnel


- Next Destination-Currently Detector testing and communications with Ministry of Land, Infrastructure and Transport
- World 5th longest undersea tunnel
- West Sea Tide : One of the largest in the world (~10m max)
- Tsunami Early Warning system using HAWL



□ 보령 해저터널 제원

○ 연 장: L=6,927m(해저구간 5,209m), B=8.5m(2차로 병렬)


수 심 : 평균 해수면에서 터널상단까지 최대 80m




□ 완공 조감도



Similarly, we mapped undersea tunnel with moving muon detector




Charge spectrum of two panels



Long-term measurements at the location where it shows low overburden (high flux) but variable water height.

Tidal Level Measurement


Model based on water level data measured at the nearby KHOA station






Chung-Ang University Neutrino Detection Yolk

50 cm




One motivation for position sensitive detector


- · A closer position at the ground level.
- If we can put CANDY to the dome, we can have a similar baseline as NEOS.
- Now, radial direction has no constraints
- CANDY can be placed in radial direction on a container box.
- Muon rate needs measurements but can be dealt with additional shielding




Position-sensitive SiPM array: Simulations and Reconstruction



Identification of positron scintillation




Using NEOS skeleton, extend to the radial direction (One more Target LS barrel)




5 mm PTFE reflector

CANDY in a Borated Water Shield for muon and neutron tagging







mini-CANDY (4 SiPMs, two each side)




Data Acquisition Setup

8ch. Coincident PS + LS data. 62.5 MSPS FADC DAQ, 56.5 V






Position-sensitive SiPM array: Data




Gamma Background Measurements



Position-sensitive detector can be made with a larger array

 $50 \times 50 \times 50$ cm Acrylic Cube with Plates and SiPMs

Medium-CANDY ongoing












SiPM Array (One layer detector with 25 SiPMs, LAB-LS ~10 cm thick)



Medium-CANDY v1 data (Muon Candidate)



Medium-CANDY v1 data (Muon-decay Candidate)



Summary and Outlook

We focus on position sensitive scintillation detectors

- We focus on R&Ds for position-sensitive scintillation detectors
- •HAWL project turns out to be very useful and we'd like to improve as much as we can.
- •Our goal is to achieve sub-centimeter resolution scintillation detector (See Minsu's talk later)
- Currently building Medium CANDY (50cmX50cmX50cm, 125 SiPM) with hardware costs roughly at 20k usd.

Rare Event Search Laboratory (RESL)

